

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVALYT® 2 – 11	Kat.Nr./Cat.No. : 42900
	Lot/Contr.No. : P250381

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Klare Lösung von Polyamino-polysulfon-polycarboxylsäuren, 40% (w/v) in Wasser, sterilfiltriert Clear solution of polyamino-polysulfonic-polycarboxylic acids, 40% (w/v) in water, sterile filtered		
pH	2% in H ₂ O, RT	6.85 ± 0.5	7.0
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.350 – 1.200	0.771
pH-Bereich pH-range	IEF ⁽¹⁾ , 150 µm-Gel, 5 °C, 3 500 Vh	2.8 ± 0.6 – 10.4 ± 0.8	3.2 – 9.6
ausgebildeter pH-Gradient (%) pH-gradient formed	IEF, 150 µm-Gel, 5 °C, 3 500 Vh	min. 80	81.2
Mindesthaltbarkeit minimum shelf life			17.11.2030
Lagerung (°C) storage			+2 bis/ to +8

(1) Gelelektrophorese wird auf Standard-Polyacrylamidgelen durchgeführt.
Gel electrophoresis is performed using a standard-polyacrylamide-gel.
(5% T, 3% C, 4% Servalyt®, 0.1% (w/v) APS, 0.078% (w/v) TEMED,
standard electrode solution # 42984 and # 42986)

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.11.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.