

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS



SERVALYT® 2 – 4	Kat.Nr./Cat.No. : 42902
	Lot/Contr.No. : P220378

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Klare Lösung von Polyamino-polysulfon-polycarboxylsäuren, 40% (w/v) in Wasser, sterilfiltriert Clear solution of polyamino-polysulfonic-polycarboxylic acids, 40% (w/v) in water, sterile filtered		
pH	2% in H ₂ O, RT	3.30 ± 0.4	3.63
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.200 – 0.800	0.318
pH-Bereich pH-range	IEF ⁽¹⁾ , 150 µm-Gel, 5°C, 5 000 Vh	2.4 ± 0.5 – 4.1 ± 0.3	2.9 – 4.3
ausgebildeter pH-Gradient (%) pH-gradient formed	IEF, 150 µm-Gel, 5°C, 5 000 Vh	min. 47	50
Mindesthaltbarkeit minimum shelf life			19.09.2027
Lagerung (°C) storage			+2 bis/ to +8

¹⁾ Gelelektrophorese wird auf Standard-Polyacrylamid-Gelen durchgeführt.
Gel electrophoresis is performed using a standard-polyacrylamide-gel.
(5% T, 3% C, 4% Servalyt®, 0.1% (w/v) APS, 0.078% (w/v) TEMED,
standard electrode solution # 42984 and # 42986)

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 31.10.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.