

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVALYT[®] PRECOTES[®] pH 3 – 6	Kat.Nr./Cat.No. : 42919
150µm, 245x125mm	Lot/Contr.No. : P130217

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 245x125mm, 150µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 245x125mm, 150µm		
pH-Servalyt	2% in H ₂ O, RT	5.06 ± 0.4	4.9
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.15 – 0.60	0.234
Proteintrennung protein separation	IEF ⁽¹⁾ , 150 µm-Gel, 5°C, 5000 Vh	entspricht Standard conform reference	entspricht conform
Haftung des Gels adhesion of the gel	IEF ⁽¹⁾ , 150 µm-Gel, 5°C, 5000 Vh	Haftung ja adhesion yes	ja yes
Mindesthaltbarkeit minimum shelf life			12/ 2014
Lagerung (°C) storage			+2 bis/ to +8

⁽¹⁾ Gelelektrophorese wird auf Standard-Polyacrylamid-Gelen durchgeführt.
Gel electrophoresis is performed using a standard-polyacrylamide-gel.
(5% T, 3% C, 4% Servalyt[®], 0.1% (w/v) APS, 0.078% (w/v) TEMED,
standard electrode solution # 42984 and # 42986)

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 14.06.2013

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.