

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVALYT[®] 2 - 9	Kat.Nr./Cat.No. :	42935
	Lot/Contr.No. :	P160418

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Klare Lösung von Polyamino-polysulfon-polycarboxylsäuren, 40% (w/v) in Wasser, steriltfiltriert Clear solution of polyamino-polysulfonic-polycarboxylic acids, 40% (w/v) in water, sterile filtered		
pH	2% in H ₂ O, RT	6.4 ± 0.3	6.3
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.55 – 0.89	0.64
pH-Bereich pH-range	IEF ⁽¹⁾ , 150 µm-Gel, 5°C, 5 000 Vh	2.5 ± 0.5 – 8.7 ± 0.4	2.7 – 8.6
ausgebildeter pH-Gradient (%) pH-gradient formed	IEF, 150 µm-Gel, 5°C, 5 000 Vh	min. 70	94.9
Mindesthaltbarkeit minimum shelf life	11.11.2021		
Lagerung (°C) storage	+2 bis/ to +8		

⁽¹⁾Gelelektrophorese wird auf Standard-Polyacrylamid-Gelen durchgeführt.
Gel electrophoresis is performed using a standard-polyacrylamide-gel.
(5% T, 3% C, 4% Servalyt[®], 0.1% (w/v) APS, 0.078% (w/v) TEMED,
standard electrode solution # 42984 and # 42986)

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.11.2016

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.