

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Glasplatten	Kat.Nr./Cat.No. :	42952
Glass plates	Lot/Contr.No. :	240046

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Größe (mm) size		265 x 128 x 3	entspricht corresponds
Aussehen appearance		klare, durchsichtige Platten clear, transparent plates	entspricht corresponds
Anwendung application		Träger für Gelfolien supports for gel sheets	entspricht corresponds

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 18.01.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.