

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVALYT® PRECOTES® pH 3-6	Kat.Nr./Cat.No. :	42974
150µm, 125x125mm	Lot/Contr.No. :	P100322

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 125x125mm, 150µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 125x125mm, 150µm		
pH-Servalyt	2% in H ₂ O, RT	5.06 ± 0.4	4.87
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.15 - 0.60	0.262
Proteintrennung protein separation	IEF, 150 µm-Gel, 5°C, 5000 Vh	entspricht Standard conform reference	entspricht conform
Haftung des Gels adhesion of the gel	IEF, 150 µm-Gel, 5°C, 5000 Vh	Haftung ja adhesion yes	ja yes
Mindesthaltbarkeit minimum shelf life			03/ 2012
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 01.10.2010

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.