

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Anodenflüssigkeit 3 für IEF	Kat.Nr./Cat.No. : 42984
Anode fluid 3 for IEF	Lot/Contr.No. : P130410

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose Lösung clear, colorless solution	entspricht corresponds
Inhalt (in 50 ml) content		0.17g L-Asparaginsäure 0.18g L-Glutaminsäure 0.17g L-aspartic acid 0.18g L-glutamic acid	
pH		2.8 – 3.8	3.1
Osmolarität (mOsmol/l) osmolarity		40 – 70	42
Leitfähigkeit (µS/cm) conductivity		< 1000	524
Anwendung (IEF) use		entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			10/ 2014
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.10.2013

Dipl.-Ing. (FH) Bernhard Göckel

Jakob Arndt

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.