

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Elektrodenochte Standardgröße	Kat.Nr./Cat.No. : 42988
Electrode wicks standard size, Filter cardboard	Lot/Contr.No. : 200665

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Größe size		120 x 6 x 1mm	entspricht corresponds
Anwendung application		für Isoelektrische Fokussierung for Isoelectric focussing	
Mindesthaltbarkeit minimum shelf life			29.06.2025
Lagerung (°C) storage			+15 bis/ to +30

SERVA Electrophoresis GmbH

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.06.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.