

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Applikatorstreifen 7 x 1	Kat.Nr./Cat.No. : 42989
Applicator strips 7 x 1	Lot/Contr.No. : 230686

Parameter parameter	Spezifikation specification	Ergebnis result
Länge length	263 x 6 x 1 mm min. 24 Probenschlitze/ sample slots: 7 x 1 mm	
Anwendung application	Probenauftrag sample application	
Mindesthaltbarkeit minimum shelf life		16.06.2028
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 20.06.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.