

ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS



---

<b>IPG Strip 4 - 7 / 7 cm</b>	<b>Kat.Nr./Cat.No.:</b>	<b>43003</b>
	<b>Lot/Contr.No. :</b>	<b>IPG24012</b>

---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Beschreibung</b> description	Getrocknete Gelstreifen mit immobilisiertem pH-Gradienten. Die homogene Acrylamidmatrix ist an die Trägerfolie GEL-FIX™ kovalent gebunden, um das Gel zu stabilisieren. Dried gel strips with immobilized pH-gradient. The homogeneous acrylamide matrix is covalently bound to the support film GEL-FIX™ to stabilize the gel.		
<b>Proteintrennung</b> protein separation	Isoelektische Fokussierung/ Isoelectric focussing		entspricht/ conform
<b>Mindesthaltbarkeit</b> minimum shelf life			25.04.2026
<b>Lagerung (°C)</b> storage			-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

---

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 10.05.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.