

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

SERVAGel (TM) Neutral HSE
10 cm x 10 cm, 1 mm

Kat.Nr./Cat.No. : 43245
Lot/Contr.No. : V180057-1

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm		
description	Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x 10 cm, Thickness: 1 mm		
Testbedingungen	Elektrophorese des SERVA Unstained SDS PAGE Protein Markers in 1x SDS Tris-Glycin (Laemmli)-Puffer bei 400 V und 20 min unter „High Speed Elektrophorese“ Bedingungen bis Lauffront Gelende erreicht.		
testing conditions	Electrophoresis of SERVA Unstained SDS PAGE Protein Marker in 1x SDS Tris-Glycine (Laemmli)-buffer at 400 V and 20 min under „high speed electrophoresis“ conditions until running front has reached end of gel.		
Proteintrennung	Myosin	200 kDa	entspricht/conform
protein separation	Ovalbumin	45 kDa	entspricht/conform
	Aprotinin	6.5 kDa	entspricht/conform
Aussehen	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.		
appearance	Protein separation shows sharp and straight bands over the whole separation range of the gel.		
Mindesthaltbarkeit	27.06.2019		
minimum shelf life			
Lagerung (°C)	+2 bis/ to +8		
storage			

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.03.2018

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.