

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS



SERVAGel(TM) Neutral HSE
10 cm x 10 cm, 1 mm

Kat.Nr./Cat.No. : 43246
Lot/Contr.No. : V23051-1

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm		
Testbedingungen testing conditions	Elektrophorese des SERVA Protein-Testmischung 6 für SDS PAGE in 1x SDS Tris-Glycin (Laemmli)-Puffer bei 400 V und 20 min unter „High Speed Elektrophorese“ Bedingungen bis Lauffront Gelende erreicht. Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE in 1x SDS Tris-Glycine (Laemmli)-buffer at 400 V and 20 min under „high speed electrophoresis“ conditions until running front has reached end of gel.		
Proteintrennung protein separation	Albumin (BSA) 67 kDa Ovalbumin 45 kDa Trypsin Inhibitor 21 kDa (soy bean)	entspricht/conform entspricht/conform entspricht/conform	entspricht/conform entspricht/conform entspricht/conform
Aussehen appearance	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels. Protein separation shows sharp and straight bands over the whole separation range of the gel.		
Mindesthaltbarkeit minimum shelf life	20.07.2024		
Lagerung (°C) storage	+2 bis/ to +8		

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 21.04.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.