

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>SERVAGel(TM) TG PRiME(TM) 10 %</b>	<b>Kat.Nr./Cat.No. : 43264</b>
<b>10 cm x 10 cm, 1 mm, 10 sample wells</b>	<b>Lot/Contr.No. : V170035-1</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 10 Probetaschen  Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 10 sample wells		
<b>Testbedingungen</b> testing conditions	Elektrophorese des SERVA Unstained SDS PAGE Protein Markers in 1x Laemmli Puffer unter Standardbedingungen bis Lauffront Gelende erreicht.  Electrophoresis of SERVA Unstained SDS PAGE Protein Marker in 1x Laemmli buffer under standard conditions until running front has reached end of gel.		
<b>Proteintrennung</b> protein separation	β-Galactosidase	116 kDa	entspricht/conform entspricht/conform
	Ovalbumin	45 kDa	entspricht/conform entspricht/conform
	Carbonic anhydrase	29 kDa	entspricht/conform entspricht/conform
<b>Aussehen</b> appearance	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.  Protein separation shows sharp and straight bands over the whole separation range of the gel.		
<b>Mindesthaltbarkeit</b> minimum shelf life			23.03.2018
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

<b>SERVA Electrophoresis GmbH</b>	<b>Druckdatum/ printing date: 07.03.2017</b>
<b>Qualitätskontrolle/ Quality Control</b>	

Dipl.-Ing. (FH) Bernhard Göckel Marina Konter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.