

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS



SERVAGel(TM) TG PRiME(TM) 14 %
10 cm x 10 cm, 1 mm, 12 sample wells

Kat.Nr./Cat.No. : 43269
Lot/Contr.No. : V200167-1

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 12 Probetaschen		
description	Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 12 sample wells		
Testbedingungen	Elektrophorese des SERVA Unstained SDS PAGE Protein Markers in 1x Laemmli Puffer unter Standardbedingungen bis Lauffront Gelende erreicht.		
testing conditions	Electrophoresis of SERVA Unstained SDS PAGE Protein Marker in 1x Laemmli buffer under standard conditions until running front has reached end of gel.		
Proteintrennung protein separation	β-Galactosidase	116 kDa	entspricht/conform
	Ovalbumin	45 kDa	entspricht/conform
	Carbonic anhydrase	29 kDa	entspricht/conform
Aussehen	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.		
appearance	Protein separation shows sharp and straight bands over the whole separation range of the gel.		
Mindesthaltbarkeit minimum shelf life	27.12.2021		
Lagerung (°C) storage	+2 bis/ to +8		

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 30.11.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.