

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS



SERVAGel(TM) TG PRiME(TM) 12 %	Kat.Nr./Cat.No. : 43286
10 cm x 10 cm, 1 mm, 15 sample wells	Lot/Contr.No. : V23021-1

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 15 Probetaschen		
description	Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 15 sample wells		
Testbedingungen	Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.		
testing conditions	Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.		
Proteintrennung	Albumin (BSA)	67 kDa	entspricht/conform
protein separation	Ovalbumin	45 kDa	entspricht/conform
	Carbonic anhydrase	29 kDa	entspricht/conform
Aussehen	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.		
appearance	Protein separation shows sharp and straight bands over the whole separation range of the gel.		
Mindesthaltbarkeit	14.03.2024		
minimum shelf life			
Lagerung (°C)	+2 bis/ to +8		
storage			

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 16.02.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.