

**A N A L Y S E N Z E R T I F I K A T**

CERTIFICATE OF ANALYSIS

<b>SERVAGel(TM) TG PRiME(TM) 8 - 16 % 10 cm x 10 cm, 1 mm, 15 sample wells</b>	<b>Kat.Nr./Cat.No. : 43290 Lot/Contr.No. : V160162-1</b>
--	--

<b>Parameter parameter</b>	<b>Methode method</b>	<b>Spezifikation specification</b>	<b>Ergebnis result</b>
<b>Beschreibung</b>	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 15 Probestaschen		
description	Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 15 sample wells		
<b>Testbedingungen</b>	Elektrophorese des SERVA Unstained SDS PAGE Protein Markers in 1x Laemmli Puffer unter Standardbedingungen bis Lauffront Gelende erreicht.		
testing conditions	Electrophoresis of SERVA Unstained SDS PAGE Protein Marker in 1x Laemmli buffer under standard conditions until running front has reached end of gel.		
<b>Proteintrennung</b>	Myosin	200 kDa	entspricht/conform
protein separation	Ovalbumin	45 kDa	entspricht/conform
	Aprotinin	6.5 kDa	entspricht/conform
<b>Aussehen</b>	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.		
appearance	Protein separation shows sharp and straight bands over the whole separation range of the gel.		
<b>Mindesthaltbarkeit</b>			03.06.2017
minimum shelf life			
<b>Lagerung (°C)</b>			+2 bis/ to +8
storage			

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH  
Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 07.11.2016**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.