

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>FocusGel 6-11 24S</b>	<b>Kat.Nr./Cat.No.:</b> 43329
<b>Size: 250 mm x 115 mm x 0.65 mm</b>	<b>Lot/Contr.No.:</b> 170593

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamid-Gele mit Träger-Ampholyt, die an eine Kunststoffolie gebunden sind. Format: 250 x 115 mm, Dicke: 0.65 mm  Ready-to-use polyacrylamide-gels with carrier ampholyte, bound to a plastic foil. Format: 250 x 115 mm, Thickness: 0.65 mm		
<b>pH-Ampholyt</b>		7.0 – 7.6	7.25
<b>Leitfähigkeit (mS/cm)</b> conductivity		1.20 – 1.70	1.61
<b>Proteintrennung</b> protein separation	IEF, 650 µm-Gel, 7 °C, 3500 Vh	entspricht Standard conform reference	entspricht conform
<b>Haftung des Gels</b> adhesion of gel	IEF, 650 µm-Gel, 7 °C, 3500 Vh	Haftung ja adhesion yes	ja yes
<b>Mindesthaltbarkeit</b> minimum shelf life			19.12.2017

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 23.06.2017

Dipl.-Ing. (FH) Bernhard Göckel

Marina Konter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.