

**ANALYSENZERTIFIKAT**  
 CERTIFICATE OF ANALYSIS

<b>FocusGel 6-11 24S</b>	<b>Kat.Nr./Cat.No. : 43329</b>
<b>Size: 250 mm x 115 mm x 0.65 mm</b>	<b>Lot/Contr.No. : 211012</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamid-Gele mit Träger-Ampholyt, die an eine Kunststoffolie gebunden sind. Format: 250 x 115 mm, Dicke: 0.65 mm  Ready-to-use polyacrylamide-gels with carrier ampholyte, bound to a plastic foil. Format: 250 x 115 mm, Thickness: 0.65 mm		
<b>pH-Ampholyt</b>		7.0 – 7.6	7.25
<b>Leitfähigkeit (mS/cm)</b> conductivity		1.20 – 1.70	1.54
<b>Proteintrennung</b> protein separation	IEF, 650 µm-Gel, 7 °C, 3500 Vh	entspricht Standard conform reference	entspricht conform
<b>Haftung des Gels</b> adhesion of gel	IEF, 650 µm-Gel, 7 °C, 3500 Vh	Haftung ja adhesion yes	ja yes
<b>Mindesthaltbarkeit</b> minimum shelf life			24.02.2022

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date: 03.09.2021**

Dr. Judith Koch

Marina Konter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.