

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

FocusGel 6-9 24S HEM

Kat.Nr./Cat.No. : 43330

Size: 250 mm x 115 mm x 0.65 mm

Lot/Contr.No. : 221098

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamid-Gele mit Träger-Ampholyt, die an eine Kunststoffolie gebunden sind. Format: 250 x 115 mm, Dicke: 0.65 mm Ready-to-use polyacrylamide-gels with carrier ampholyte, bound to a plastic foil. Format: 250 x 115 mm, Thickness: 0.65 mm		
pH-Ampholyt		6.6 – 7.6	7.26
Leitfähigkeit (mS/cm) conductivity		0.20 – 1.00	0.72
Proteintrennung protein separation	IEF, 650 µm-Gel, 7 °C, 4250 Vh	entspricht Standard conform reference	entspricht conform
Haftung des Gels adhesion of gel	IEF, 650 µm-Gel, 7 °C, 4250 Vh	Haftung ja adhesion yes	ja yes
Mindesthaltbarkeit minimum shelf life			08.11.2023
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.11.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.