

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

CleanGel IEF for PhastSystem	Kat.Nr./Cat.No. : 43350
Size: 50 mm x 42 mm x 0.43 mm	Lot/Contr.No. : 150068

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	20 rehydratisierbare Polyacrylamid-Gele kovalent an eine Kunststoffolie gebunden, 5 % T, ohne Probetaschen. Format: 50 x 42 mm, Dicke: 0.43 mm		
	20 rehydratable polyacrylamide-gels covalently bound to a plastic sheet, 5 % T, without slots. Format: 50 x 42 mm, Thickness: 0.43 mm		
Mindesthaltbarkeit minimum shelf life			07.10.2017
Lagerung (°C) storage			-15 bis/ to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.09.2016

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.