

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>CleanGel IEF for PhastSystem</b>	<b>Kat.Nr./Cat.No. : 43350</b>
Size: 50 mm x 42 mm x 0.43 mm	<b>Lot/Contr.No. : 221002</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Beschreibung</b>	20 rehydratisierbare Polyacrylamid-Gele kovalent an eine Kunststoffolie gebunden, 5 % T, ohne Probetaschen. Format: 50 x 42 mm, Dicke: 0.43 mm		
description	20 rehydratable polyacrylamide-gels covalently bound to a plastic sheet, 5 % T, without slots. Format: 50 x 42 mm, Thickness: 0.43 mm		
<b>Mindesthaltbarkeit</b>			23.08.2024
minimum shelf life			
<b>Lagerung (°C)</b>			-15 bis/ to -25
storage			

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 26.10.2022**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.