

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

CleanGel IEF for PhastSystem	Kat.Nr./Cat.No.:	43350
Size: 50 mm x 42 mm x 0.43 mm	Lot/Contr.No.:	230926

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung	20 rehydratisierbare Polyacrylamid-Gele kovalent an eine Kunststofffolie gebunden, 5 % T, ohne Probetaschen. Format: 50 x 42 mm, Dicke: 0.43 mm		
description	20 rehydratable polyacrylamide-gels covalently bound to a plastic sheet, 5 % T, without slots. Format: 50 x 42 mm, Thickness: 0.43 mm		
Mindesthaltbarkeit minimum shelf life			09.01.2025
Lagerung (°C) storage			-15 bis/ to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 12.09.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.