

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA HPE™ IPG strip buffer	Kat.Nr./Cat.No. : 43368
	Lot/Contr.No. : P230018

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Klare Lösung von Polyamino-polysulfon- polycarboxylsäuren, 40% (w/v) in Wasser, steril filtriert Clear solution of polyamino-polysulfonic-polycarboxylic acids, 40% (w/v) in water, sterile filtered		
pH	2 % in H ₂ O, RT	6.6 ± 0.4	6.6
Leitfähigkeit (mS/cm) conductivity	2 % in H ₂ O, RT	0.300 – 0.700	0.505
Anwendung application	2D Elektrophorese 2D electrophoresis	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life	05.05.2027		
Lagerung (°C) storage	+2 bis/ to +8		

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 12.01.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.