

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Cooling Contact Fluid	Kat.Nr./Cat.No.:	43371
	Lot/Contr.No.:	P220402

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Anwendung application	für den Gebrauch in der horizontalen Elektrophorese for use in horizontal electrophoresis		
Mindesthaltbarkeit minimum shelf life			16.11.2025
Lagerung (°C) storage			+2 bis/ to +8

SERVA Electrophoresis GmbH

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 21.11.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.