

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

SERVA Purple

Kat.Nr./Cat.No. : 43386

Lot/Contr.No. : P240089

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, gelbe Lösung clear, yellow solution	entspricht corresponds
Proteindetektion protein detection	10x in H ₂ O mit Albumin 10x in H ₂ O with albumin	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			11.03.2026
Lagerung (°C) storage			-15 bis/ to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 11.04.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.