

**ANALYSENZERTIFIKAT**  
 CERTIFICATE OF ANALYSIS

|  |                                |
|--|--------------------------------|
| <b>SDS Gels 10% 25S</b>                | <b>Kat.Nr./Cat.No. : 43850</b> |
| <b>Size: 250 mm x 125 mm x 0.45 mm</b> | <b>Lot/Contr.No. : 180696</b>  |

| Parameter<br>parameter    | Methode<br>method   | Spezifikation<br>specification | Ergebnis<br>result |
|---------------------------|---|--------------------------------|--------------------|
| <b>Beschreibung</b>       | Gebrauchsfertige Polyacrylamid-Gele, die an eine Kunststoffolie gebunden sind, zur Verwendung in der SDS-Gelelektrophorese von Proteinen.<br>Format: 250 x 125 mm, Dicke: 0.45 mm |                                |                    |
| description               | Ready-to-use polyacrylamide-gels, bound to a plastic foil, for use in SDS gel electrophoresis of proteins.<br>Format: 250 x 125 mm, Thickness: 0.45 mm                            |                                |                    |
| <b>Testbedingungen</b>    | Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.  |                                |                    |
| testing conditions        | Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.  |                                |                    |
| <b>Proteintrennung</b>    | SDS-Gelelektrophorese   | entspricht                     | entspricht         |
| protein separation        | SDS gel electrophoresis   | conforms                       | conforms           |
| <b>Mindesthaltbarkeit</b> |   |                                | 21.06.2020         |
| minimum shelf life        |   |                                |                    |
| <b>Lagerung (°C)</b>      |   |                                | +2 bis/ to +8      |
| storage                   |   |                                |                    |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date: 02.07.2018**

Dr. Judith Koch

Marina Konter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                      |                         |               |
|--------------------------------------|-------------------------|---------------|
| <b>Buffer Kit for SDS Gels Urine</b> | <b>Kat.Nr./Cat.No.:</b> | <b>43385</b>  |
|                                      | <b>Lot/Contr.No.:</b>   | <b>180666</b> |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method  | <b>Spezifikation</b><br>specification | <b>Ergebnis</b><br>result |
|---|---|---------------------------------------|---------------------------|
| <b>Beschreibung</b><br>description              | Kit enthält SDS-Anodenpuffer, SDS-Kathodenpuffer, SDS-Probenverdünnungspuffer neutral 10X Urin und SDS-Probenverdünnungspuffer neutral 1X sowie Kühlkontaktflüssigkeit.<br>Kit contains SDS anode buffer, SDS cathode buffer and SDS buffer neutral sample diluter 10X urine, SDS buffer neutral sample diluter 1X and cooling contact fluid. |                                       |                           |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |   |                                       | 23.10.2019                |
| <b>Lagerung (°C)</b><br>storage                 |   |                                       | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 19.06.2018

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.