

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

SDS Gels 10% 25S	Kat.Nr./Cat.No. : 43850
Size: 250 mm x 125 mm x 0.45 mm	Lot/Contr.No. : 210898

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung	Gebrauchsfertige Polyacrylamid-Gele, die an eine Kunststoffolie gebunden sind, zur Verwendung in der SDS-Gelelektrophorese von Proteinen. Format: 250 x 125 mm, Dicke: 0.45 mm		
description	Ready-to-use polyacrylamide-gels, bound to a plastic foil, for use in SDS gel electrophoresis of proteins. Format: 250 x 125 mm, Thickness: 0.45 mm		
Testbedingungen	Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.		
testing conditions	Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.		
Proteintrennung protein separation	SDS-Gelelektrophorese SDS gel electrophoresis	entspricht conforms	entspricht conforms
Mindesthaltbarkeit minimum shelf life			26.07.2023
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.07.2021

Dr. Judith Koch

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Buffer Kit for SDS Gels Urine	Kat.Nr./Cat.No.:	43385
	Lot/Contr.No.:	210882

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Kit enthält SDS-Anodenpuffer, SDS-Kathodenpuffer, SDS-Probenverdünnungspuffer neutral 10X Urin und SDS-Probenverdünnungspuffer neutral 1X sowie Kühlkontaktflüssigkeit. Kit contains SDS anode buffer, SDS cathode buffer and SDS buffer neutral sample diluter 10X urine, SDS buffer neutral sample diluter 1X and cooling contact fluid.		
Mindesthaltbarkeit minimum shelf life			26.04.2023
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 02.08.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.