

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

| | |
|--|--------------------------------|
| SDS Gels 10% 25S | Kat.Nr./Cat.No. : 43850 |
| Size: 250 mm x 125 mm x 0.45 mm | Lot/Contr.No. : 230633 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|---|---------------------------------------|---------------------------|
| Beschreibung | Gebrauchsfertige Polyacrylamid-Gele, die an eine Kunststoffolie gebunden sind, zur Verwendung in der SDS-Gelelektrophorese von Proteinen. Format: 250 x 125 mm, Dicke: 0.45 mm | | |
| description | Ready-to-use polyacrylamide-gels, bound to a plastic foil, for use in SDS gel electrophoresis of proteins. Format: 250 x 125 mm, Thickness: 0.45 mm | | |
| Testbedingungen | Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht. | | |
| testing conditions | Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel. | | |
| Proteintrennung protein separation | SDS-Gelelektrophorese SDS gel electrophoresis | entspricht conforms | entspricht conforms |
| Mindesthaltbarkeit minimum shelf life | | | 26.05.2025 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 02.06.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|--------------------------------------|-------------------------|---------------|
| Buffer Kit for SDS Gels Urine | Kat.Nr./Cat.No.: | 43385 |
| | Lot/Contr.No.: | 230252 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|---|---------------------------------------|---------------------------|
| Beschreibung description | Kit enthält SDS-Anodenpuffer, SDS-Kathodenpuffer, SDS-Probenverdünnungspuffer neutral 10X Urin und SDS-Probenverdünnungspuffer neutral 1X sowie Kühlkontaktflüssigkeit. Kit contains SDS anode buffer, SDS cathode buffer and SDS buffer neutral sample diluter 10X urine, SDS buffer neutral sample diluter 1X and cooling contact fluid. | | |
| Mindesthaltbarkeit minimum shelf life | | | 25.07.2024 |
| Lagerung (°C) storage | | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 03.03.2023

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.