

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Anti-Corrosive Additive	Kat.Nr./Cat.No.:	43392
4x concentrate	Lot/Contr.No.:	160816

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, rosa Lösung clear, pink solution	entspricht corresponds
pH		8.2 – 8.6	entspricht corresponds
Dichte density	20 °C	1.122 – 1.125	entspricht corresponds
Brechungsindex refractive index	20 °C	1.432 – 1.436	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			24.08.2021
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.08.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.