

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                |                         |               |
|--------------------------------|-------------------------|---------------|
| <b>Anti-Corrosive Additive</b> | <b>Kat.Nr./Cat.No.:</b> | <b>43392</b>  |
| 10x concentrate                | <b>Lot/Contr.No.:</b>   | <b>220749</b> |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method | <b>Spezifikation</b><br>specification      | <b>Ergebnis</b><br>result |
|-------------------------------------------------|--------------------------|--------------------------------------------|---------------------------|
| <b>Aussehen</b><br>appearance                   |                          | klare, rosa Lösung<br>clear, pink solution | entspricht<br>corresponds |
| <b>pH</b>                                       |                          | 8.2 – 8.6                                  | entspricht<br>corresponds |
| <b>Dichte</b><br>density                        | 20 °C                    | 1.122 – 1.125                              | entspricht<br>corresponds |
| <b>Brechungsindex</b><br>refractive index       | 20 °C                    | 1.432 – 1.436                              | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                          |                                            | 29.07.2027                |
| <b>Lagerung (°C)</b><br>storage                 |                          |                                            | +15 bis/ to +30           |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 01.08.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.  
This report has been computer generated and does not contain a signature.