

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS



SERVA HPE(TM) Silver Staining Kit

Kat.Nr./Cat.No. : 43395

Lot/Contr.No. : 240269

Parameter
parameter

Spezifikation
specification

Ergebnis
result

Inhalt
content

Kit enthält Reagenzien zum Färben von
4 Large 2D HPE™ Gelen.
Kit contains reagents to stain 4 large 2D
HPE™ gels.

Mindesthaltbarkeit
minimum shelf life

31.03.2025

Lagerung (°C)
storage

+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 21.03.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.