

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA HPE IPG Overlay	Kat.Nr./Cat.No. :	43397
	Lot/Contr.No. :	121255

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Öl zur Überschichtung von IPG-Streifen während der Isoelektrischen Fokussierung Overlay oil to cover IPG strips during isoelectric focusing		
Aussehen appearance		klare, farblose, viskose Flüssigkeit clear, colorless, viscous liquid	entspricht corresponds
Brechungsindex refractive index	n_D^{20}	1.468 – 1.470	1.469
Mindesthaltbarkeit minimum shelf life			08/ 2017
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 08.11.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.