

ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

---

<b>SERVA HPE IPG Overlay</b>	<b>Kat.Nr./Cat.No. :</b>	<b>43397</b>
	<b>Lot/Contr.No. :</b>	<b>130481</b>

---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Beschreibung</b> description		Öl zur Überschichtung von IPG-Streifen während der Isoelektrischen Fokussierung Overlay oil to cover IPG strips during isoelectric focusing	
<b>Aussehen</b> appearance		klare, farblose, viskose Flüssigkeit clear, colorless, viscous liquid	entspricht corresponds
<b>Brechungsindex</b> refractive index	$n_D^{20}$	1.468 – 1.470	1.469
<b>Mindesthaltbarkeit</b> minimum shelf life			08/ 2017
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

---

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 17.04.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.