

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVA Lightning Sci3	Kat.Nr./Cat.No. :	43405
	Lot/Contr.No. :	170529

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
monoisotope Masse (Da) monoisotopic mass		582.33	
Aussehen appearance		rötliches Lyophilisat reddish lyophilisate	entspricht corresponds
Anregung (nm) excitation		555	
Emission (nm)		570	
Anwendungstest application test	für 2D PAGE for 2D PAGE	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			04.12.2019
Lagerung (°C) storage			-15 bis / to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 07.12.2018

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.