

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>SERVA Lightning Sci5</b>	<b>Kat.Nr./Cat.No. :</b>	<b>43406</b>
	<b>Lot/Contr.No. :</b>	<b>201062</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>monoisotope Masse (Da)</b> monoisotopic mass		580.32	
<b>Aussehen</b> appearance		blaues Pulver blue powder	entspricht corresponds
<b>Anregung (nm)</b> excitation		645	
<b>Emission (nm)</b>		660	
<b>Anwendungstest</b> application test	für 2D PAGE for 2D PAGE	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			06.10.2021
<b>Lagerung (°C)</b> storage			-15 bis / to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 07.10.2020**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.