

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>SERVA SDS Gel 8-25 % Kit for PhastSystem</b>	<b>Kat.-Nr./Cat. No.:</b>	<b>43503</b>
Size: 50 mm x 42 mm	<b>Lot/Contr.No. :</b>	<b>220579</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Beschreibung</b>	10 gebrauchsfertige SDS-Polyacrylamid-Gele kovalent an eine Kunststoffolie gebunden (Format: 50 x 42 mm) 20 Elektroden-Wicks SDS-Kathoden- und SDS-Anoden-Puffer (je 20 ml)		
description	10 ready-to-use SDS polyacrylamide gels covalently bound to a plastic sheet (size: 50 x 42 mm) 20 electrode wicks SDS cathode and SDS anode buffer (20 ml each)		
<b>Mindesthaltbarkeit</b> minimum shelf life			27.04.2023
<b>Lagerung (°C)</b> storage			+ 2 bis/ to + 8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 09.06.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.