

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                   |                          |               |
|-----------------------------------|--------------------------|---------------|
| <b>Visking® Dialysierschlauch</b> | <b>Kat.Nr./Cat.No. :</b> | <b>44126</b>  |
| Visking® dialysis tubing          | <b>Lot/Contr.No. :</b>   | <b>250981</b> |

| <b>Parameter</b><br>parameter                       | <b>Spezifikation</b><br>specification           | <b>Ergebnis</b><br>result |
|-----------------------------------------------------|-------------------------------------------------|---------------------------|
| <b>Material</b>                                     | regenerierte Cellulose<br>regenerated cellulose | entspricht<br>corresponds |
| <b>Flachbreite (mm)</b><br>flat width               | 77                                              | entspricht<br>corresponds |
| <b>Durchmesser (mm)</b><br>diameter                 | 49                                              | entspricht<br>corresponds |
| <b>Porendurchmesser (Å )</b><br>pore diameter       | ca. 25                                          | entspricht<br>corresponds |
| <b>Ausschlußvolumen (Dalton)</b><br>exclusion limit | 12 000 – 14 000                                 | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life     |                                                 | 20.10.2030                |
| <b>Lagerung (°C)</b><br>storage                     |                                                 | +15 bis/ to +30           |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 21.10.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.  
This report has been computer generated and does not contain a signature.