

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Visking® Dialysierschlauch</b>	<b>Kat.Nr./Cat.No. :</b>	<b>44130</b>
Visking® dialysis tubing	<b>Lot/Contr.No. :</b>	<b>130824</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Material</b>	regenerierte Cellulose regenerated cellulose	entspricht corresponds
<b>Flachbreite (mm)</b> flat width	117	entspricht corresponds
<b>Durchmesser (mm)</b> diameter	75	entspricht corresponds
<b>Porendurchmesser (Å )</b> pore diameter	ca. 25	entspricht corresponds
<b>Ausschlußvolumen (Dalton)</b> exclusion limit	12 000 – 14 000	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life		07/ 2018
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 18.07.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Jakob Arndt

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.