

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

| | |
|-----------------------------------|--------------------------------|
| Visking® Dialysierschlauch | Kat.Nr./Cat.No. : 44130 |
| Visking® dialysis tubing | Lot/Contr.No. : 190617 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---|---------------------------|
| Material | regenerierte Cellulose regenerated cellulose | entspricht corresponds |
| Flachbreite (mm) flat width | 117 | entspricht corresponds |
| Durchmesser (mm) diameter | 75 | entspricht corresponds |
| Porendurchmesser (Å) pore diameter | ca. 25 | entspricht corresponds |
| Ausschlußvolumen (Dalton) exclusion limit | 12 000 – 14 000 | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 18.06.2024 |
| Lagerung (°C) storage | | +15 bis/ to +30 |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 18.06.2019

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.