

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Servapor® Dialysierschlauch	Kat.Nr./Cat.No. :	44145
Servapor® dialysis tubing	Lot/Contr.No. :	210849

Parameter parameter	Spezifikation specification	Ergebnis result
Flachbreite (mm) flat width	ca. 25	entspricht corresponds
Durchmesser (mm) diameter	16	entspricht corresponds
Ausschlußgrenze (Dalton) exclusion limit	12 000 – 14 000	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		06.07.2026
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 14.07.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.