

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Servapor® Dialysierschlauch	Kat.Nr./Cat.No. :	44146
Servapor® dialysis tubing	Lot/Contr.No. :	160169

Parameter parameter	Spezifikation specification	Ergebnis result
Durchmesser (mm) diameter	29	entspricht corresponds
Breite flach (mm) flat width	ca. 45	entspricht corresponds
Ausschlußgrenze (MWCO) exclusion limit	12 000 – 14 000	entspricht corresponds
Porengröße (Å) pore size	ca. 25	entspricht corresponds
Lagerung (°C) storage		+15 bis/ to +30
Mindesthaltbarkeit minimum shelf life		05.11.2020

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 12.08.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.