

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                                          |                          |               |
|----------------------------------------------------------|--------------------------|---------------|
| <b>Servapor® Dialysierschlauch, MWCO 12 000 - 14 000</b> | <b>Kat.Nr./Cat.No. :</b> | <b>44148</b>  |
| Servapor® dialysis tubing, MWCO 12 000 - 14 000          | <b>Lot/Contr.No. :</b>   | <b>250518</b> |

| Parameter<br>parameter                              | Spezifikation<br>specification                  | Ergebnis<br>result        |
|-----------------------------------------------------|-------------------------------------------------|---------------------------|
| <b>Material</b><br>material                         | regenerierte Cellulose<br>regenerated cellulose | entspricht<br>corresponds |
| <b>Flachbreite (mm)</b><br>Nominal dry flat width   | 80                                              | entspricht<br>corresponds |
| <b>Durchmesser (mm)</b><br>Nominal dry diameter     | 50                                              | entspricht<br>corresponds |
| <b>Porendurchmesser (Å)</b><br>pore diameter        | ca. 25                                          | entspricht<br>corresponds |
| <b>Ausschlußvolumen (Dalton)</b><br>exclusion limit | 12 000 – 14 000                                 | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life     |                                                 | 31.07.2029                |
| <b>Lagerung (°C)</b><br>storage                     |                                                 | +15 bis/ to +30           |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 11.06.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.  
This report has been computer generated and does not contain a signature.