

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Servapor® 6 Dialysierschlauch, MWCO 6 000 - 8 000	Kat.Nr./Cat.No. :	44562
Servapor® 6 dialysis tubing, MWCO 6 000 - 8 000	Lot/Contr.No. :	210810

Parameter parameter	Spezifikation specification	Ergebnis result
Material material	regenerierte Cellulose regenerated cellulose	entspricht corresponds
Nominale Trocken-Flachbreite (mm) Nominal dry flat width	34	entspricht corresponds
Nominaler Trockendurchmesser (mm) Nominal dry diameter	22	entspricht corresponds
Ausschlußvolumen (Dalton) exclusion limit	6 000 – 8 000	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		02.07.2026
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 26.10.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.