

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Serdolit<sup>®</sup> Blau p.A.</b>	<b>Kat.Nr./Cat.No. : 45520</b>
Serdolit <sup>®</sup> Blue analytical grade	<b>Lot/Contr.No. : 160604</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		blaues Korn blue particles	entspricht corresponds
<b>Korngröße (mesh)</b> particle size		16 – 50 (0,3 – 1.2mm)	entspricht corresponds
<b>Kapazität (eq/l)</b> capacity		min. 0.8	1.04
<b>Trocknungsverlust (%)</b> loss on drying	3h, 105°C	60 – 70	63.5
<b>Gegenion</b> counter ion		OH <sup>-</sup>	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			31.05.2021
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 20.06.2016**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.