

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Methanol p.A.	Kat.Nr./Cat.No. :	45631
Methanol analytical grade	Lot/Contr.No. :	251108

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		CH ₃ OH	
Molekulargewicht molecular weight		32.04	
Aussehen appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Gehalt (%) assay	GC	min. 99.9	99.9
Brechungsindex refractive index	n _D ²⁰	1.3280 – 1.3296	1.3284
Dichte (g/ml) specific density	20 °C	0.7910 – 0.7930	entspricht corresponds
Identität identity	IR	entspricht corresponds	entspricht corresponds
Wasser (ppm) water	KF	max. 500	100
Freie Säure (meq/g) acidity		max. 0.0003	< 0.0003
Abdampfrückstand (ppm) residue on evaporation		max. 8	1
Mindesthaltbarkeit minimum shelf life			31.05.2031
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 11.11.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.