

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

Aceton	Kat.Nr./Cat.No. :	45632
Acetone research grade	Lot/Contr.No. :	211037

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₃ H ₆ O	
Molekulargewicht molecular weight		58.08	
Aussehen appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
Gehalt (%) assay	GC	min. 99.0	99.9
Dichte density	20 °C	0.790 – 0.793	0.791
Wasser (%) water	KF	max. 0.30	0.01
Mindesthaltbarkeit minimum shelf life			30.09.2025
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 07.09.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.