

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Aceton</b>	<b>Kat.Nr./Cat.No. :</b>	<b>45632</b>
Acetone research grade	<b>Lot/Contr.No. :</b>	<b>240452</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>3</sub> H <sub>6</sub> O	
<b>Molekulargewicht</b> molecular weight		58.08	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 99.0	99.9
<b>Dichte</b> density	20 °C	0.790 – 0.793	0.791
<b>Wasser (%)</b> water	KF	max. 0.30	0.02
<b>Mindesthaltbarkeit</b> minimum shelf life			30.11.2028
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 23.05.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.