

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Essigsäure p.A.</b> Glacial acetic acid p.A.	<b>Kat.Nr./Cat.No. :</b> 45633 <b>Lot/Contr.No. :</b> 170103
--	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>2</sub> H <sub>4</sub> O <sub>2</sub>	
<b>Molekulargewicht</b> molecular weight		60.05	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 99.8	≥ 99.8
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Dichte</b> density		1.05	entspricht corresponds
<b>Schwermetalle (ppm)</b> heavy metals		max. 0.5	≤ 0.5
<b>Mindesthaltbarkeit</b> minimum shelf life			31.01.2020
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 09.02.2017

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.