

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Trifluoressigsäure für LC-MS Trifluoroacetic acid for LC-MS	Kat.Nr./Cat.No. : 45641 Lot/Contr.No. : 210716
---	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		CF ₃ COOH	
Molekulargewicht molecular weight		114.02	
Aussehen appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
Gehalt (%) assay	Titration	min. 99.9	100.5
Wasser (%) water	KF	max. 0.05	0.01
Chlorid (ppm) chloride	Cl	max. 10	< 10
Fluorid (ppm) fluoride	F	max. 50	< 50
Sulfat (ppm) sulphate	SO ₄	max. 10	< 10
Mindesthaltbarkeit minimum shelf life			17.06.2024
Lagerung/storage (°C)			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.06.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.