

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Collagen R Lösung 0.2 %, steril</b>	<b>Kat.Nr./Cat.No. : 47254</b>
Collagen R solution 0.2 %, sterile	<b>Lot/Contr.No. : 121299</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Molekulargewicht</b> molecular weight		ca. 300 000	
<b>Aussehen</b> appearance		klare, viskose Lösung clear, viscous solution	entspricht corresponds
<b>Gehalt (%)</b> assay	UV	0.18 – 0.22	0.22
<b>pH</b>		3.2 – 3.8	3.6
<b>Steriltest</b> sterile test		kein Wachstum no growth	entspricht corresponds
<b>Gelbildung</b> gel formation		positiv positive	entspricht corresponds
<b>Ursprung</b> origin		Rattenschwanz rat tail	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			10/ 2014
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 27.11.2012**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.